

# COMPLIANCE

# Certificate

**Test Item:** REMANIUM<sup>®</sup> CS + \*  
**Supplier:** DENTAURUM J.P. Winkelstroeter KG  
Turnstraße 31, 75228 Ispringen, Germany

**Study performed:**

The following study (BSL Project No. 961104) was carried out in accordance with ISO 10993-5, 1992: "Tests for *in vitro* cytotoxicity"

### CYTOTOXICITY

The cytotoxic effect of the test material was examined *in vitro* by the cell culture test using L 929-cells (fibroblast cell line): Examination of materials for medical devices with the extraction method and protein determination.

**Result:** No leachable substances were released in cytotoxic concentrations (leading to a cell growth inhibition of greater than 30%) from the test item.


**BSL BIOSERVICE Scientific Laboratories GmbH**

Behringstraße 6/8

82152 Planegg

Germany

\*originally tested as REMANIUM CS<sup>®</sup>, Lot No.: 6089

  
Dr. Achim Albrecht

Head of *in vivo* Pharmacology/Toxicology

Date: 02 June 2009