


Test Certificate

Test	Cytotoxicity, L 929-Proliferation DIN EN ISO 10993-5, -12, ISO 9363-1, LM SOP 4-06-01
Test material	Cobalt chromium alloy Remanium 2000 + ¹ , order no. 102-600-01 ¹ Change of the original test material specification according to the instruction of the sponsor
Manufacturer	Dentaurum J.P. Winkelstroeter KG, Turnstrasse 31, D-75228 Ispringen
Test number	011516-20 (the complete report is available at the client)
Test result	The present results indicate that the cobalt chromium alloy Remanium 2000 + does not release cytotoxic substances.



Dr. Madlon Timme
Study Director



Dr. Stephan Rossberger