

Test Certificate

Test	Cytotoxicity, L 929-Proliferation EN ISO 10993-5, -12, EN ISO 9363-1, LM P 4-06, LM SOP 4-06-01
Test material	CoCr alloy Remanium Secura, lot 11212
Manufacturer	Dentaurum J.P. Winkelstroeter KG, Turnstrasse 31, D-75228 Ispringen
Test number	042690-20 (the complete report is available at the client)
Test result	The present results indicate that the CoCr alloy Remanium Secura does not release cytotoxic substances.



Dipl.-Ing. (FH) Jana Vogel
Study Director



Dr. Madlon Timme
Technical Supervisor
