

Test Certificate

Test Cytotoxicity, L 929-Proliferation, EN 30993-5, -12, ISO 10993-5, -12, LM SOP 4-06-01

Test material Cobalt chromium alloy Remanium GM 380 +¹
¹ Change of the original test material specification according to the instruction of the sponsor

Manufacturer Dentaurum J.P. Winkelstroeter KG, Postfach 440, D-75104 Pforzheim

Test number 000167-20 (the complete report is available at the client)

Test result The present results indicate that the cobalt chromium alloy Remanium GM 380 + does not release cytotoxic substances.



Dr. Madlon Timme
Study Director



Dr. Stephan Rossberger