

Test Certificate

| | |
|---------------|---|
| Test | Cytotoxicity, L 929-Proliferation DIN EN ISO 10993-5, -12, ISO 9363-1, LM SOP 4-06-01 |
| Test material | Cobalt chromium alloy Remanium® GM 800 + ¹ , sample pieces 15 x 10 x 1 mm ¹ Change of the original test material specification according to the instruction of the sponsor |
| Manufacturer | Dentaurum J.P. Winkelstroeter KG, Turnstrasse 31, D-75228 Ispringen |
| Test number | 020825-20 (the complete report is available at the client) |
| Test result | The present results indicate that the cobalt chromium alloy Remanium® GM 800 + does not release cytotoxic substances. |



Dr. Madlon Timme
Study Director



Dipl.-Ing. (FH) Anke Warschkow
Technical Manager