

Test Certificate

Test	Cytotoxicity, L 929-Proliferation DIN EN ISO 10993-5, -12, ISO 9363-1, LM SOP 4-06-01
Test material	CoCrW alloy Remanium Star, lot 10973
Manufacturer	Dentaurum J.P. Winkelstroeter KG, Turnstrasse 31, D-75228 Ispringen
Test number	022450-20 (the complete report is available at the client)
Test result	The present results indicate that the CoCrW alloy Remanium Star does not release cytotoxic substances.

H. Timme

Dr. Madlon Timme
Study Director

A. Warschkow

Dipl.-Ing. (FH) Anke Warschkow
Technical Manager